Review

Accuracy and reproducibility of photogrammetric measurements in dentistry: A systematic review

Oliwier Wojciech Pioterek[1,A–F], Katarzyna Majka[1,B–D], Jakub Goździewicz[1,B–D], Wiktoria Taurogińska[1,B–D], Barbara Parol[1,C,D], Daria Pozdniakowa[1,C–E], Albert Synal[1,C,D], Aleksander Sowiński[1,C,D], Martyna Piskorz[1,C,D], Mateusz Drążyk[1,D–F], Mateusz Mazurek[1,E,F], Michał Tulski[1,E], Zygmunt Domagała[2,E,F], Maciej Dobrzyński[3,E,F], Paweł Dąbrowski[2,E,F]

[1] Clinical and Dissecting Anatomy Students Scientific Club, Wroclaw Medical University, Poland
[2] Division of Anatomy, Wroclaw Medical University, Poland
[3] Department of Pediatric Dentistry and Preclinical Dentistry, Wroclaw Medical University, Poland

A – research concept and design; B – collection and/or assembly of data; C – data analysis and interpretation; D – writing the article; E – critical revision of the article; F – final approval of the article

Dental and Medical Problems, ISSN 1644-387X (print), ISSN 2300-9020 (online) *Dent Med Probl*. 2026;63(5):1313–1330

Address for correspondence
Oliwier Wojciech Pioterek
E-mail: oliwierpioterek@gmail.com

Funding sources
None declared

Conflict of interest
None declared

Acknowledgements
None declared

Received on March 12, 2025
Reviewed on May 12, 2025
Accepted on May 22, 2025

Published online on September 24, 2026

Abstract

Photogrammetry is a method of acquiring three-dimensional (3D) models by superimposing two-dimensional images. Since the publication of the last systematic review, many new articles concerning its use in dental practice have been published. This study aims to evaluate the accuracy and reproducibility of photogrammetry based on the available literature. Four databases (PubMed®, Web of Science, Google Scholar, and Embase) were searched using the following keywords: "photogrammetry"; "stereophotogrammetry"; "dentistry"; and "tooth implants". The following eligibility criteria were applied: the study had to be published in a reliable journal, specify the type of measurements performed, and describe their accuracy or reproducibility. The Quality Assessment of Diagnostic Accuracy Studies (QUADAS-1) tool was used for the assessment of the risk of bias. The collected data was tabularized and classified into 3 groups based on the type of measurements performed and into 2 subgroups according to accuracy and precision. Forty-four articles were included in the final analysis. The reviewed articles showed that cephalometric measurements based on photogrammetry have high correlation coefficients and are accurate, with measurement errors below 1 mm. Photogrammetry was highly reproducible and more accurate than intraoral scanners (IOS) for assessing implant positions. Measurements of dental arches showed high intraclass correlation coefficients (*ICC*s) and high accuracy, with errors below 0.5 mm. The main limitation of this review is that most of the included studies were conducted in vitro; therefore, it was difficult to evaluate the applicability of photogrammetry in clinical practice. Photogrammetry appears to be a precise and accurate method for performing cephalometric measurements, assessing dental implant positions and measuring dental dimensions. However, further clinical studies are needed to evaluate its practical relevance.

Keywords: dental implants, dental measurements, cephalometry, stereophotogrammetry, three-dimensional models

Cite as
Pioterek OW, Majka K, Goździewicz J, et al. Accuracy and reproducibility of photogrammetric measurements in dentistry: A systematic review. *Dent Med Probl.* 2026;63(5):1313–1330. doi:10.17219/dmp/205471

DOI
10.17219/dmp/205471

Copyright
Copyright by Author(s)
This is an article distributed under the terms of the Creative Commons Attribution 3.0 Unported (CC BY 3.0) (https://creativecommons.org/licenses/by/3.0/)

Highlights

- Photogrammetry is a highly reliable and accurate technique for dental applications, including cephalometric, dental arch and implant position measurements.
- Its effectiveness depends on external conditions, with scanning distance and light intensity identified as significant factors.
- Photogrammetry provides more precise and accurate implant position measurements than intraoral scanners and conventional impressions.
- Commercially available smartphone-based photogrammetry applications can accurately and precisely create digital three-dimensional facial models.
- Photogrammetry enables highly precise dental measurements, with errors below 0.5 mm, comparable to those of conventional techniques.

Introduction

Photogrammetry is an innovative method of obtaining three-dimensional (3D) models that is constantly being developed.[1,2] It has been used in geography and architecture,[3,4] but has now become a part of medicine.[5,6] This technique enables the creation of 3D models from two-dimensional photographs. Images of particular objects are processed using dedicated software, which superimposes photographs taken from different perspectives and creates a 3D mesh with a realistic texture based on the mathematical principle of triangulation (Fig. 1).[2] Such 3D models can later be used to perform measurements or produce 3D prints.[7–9]

The main differences between conventional scanning methods, such as intraoral and extraoral scanners, and photogrammetric methods lie in the process of capturing and processing 3D models. Intraoral scanners (IOS) mainly use laser scanning technology that irradiates an object with a laser beam, which is then reflected back to the scanner, and calculate the angle and distance between the laser, sensor and the scanned object, allowing it to generate a 3D point cloud of the object.[10] Extraoral scanners are based on structured light technology and project a geometric pattern onto the scanned object. A camera then captures the distorted pattern, on the basis of which a 3D model is generated.[11] Photogrammetry, on the other hand, is based solely on photographs captured from multiple angles, which, as mentioned above, are processed to generate a 3D model.[9]

In dentistry, conventional methods, such as IOS, laser scanners and dental impressions obtained using impression trays, are commonly used for taking measurements of the oral cavity and face.[12–15] Photogrammetry has the potential to become a useful tool for the digitization of patients' oral cavity data or an alternative to IOS for determining dental implant positions. Numerous studies have shown that it can be a less expensive alternative to IOS while being as accurate as conventional methods.[16–18] Furthermore, the use of abutments that are screwed onto a patient's implants, functioning as markers for photogrammetric machinery, allows their positions to be captured. Because the abutments are elevated above the gingiva, photogrammetry is not limited by the presence of blood or saliva, as is the case with IOS.[19,20] It is worth noting that this technology, although it has greater potential for capturing dental implant positions, is unable to capture adjacent structures, unlike IOS, which provide direct imaging of the patient's oral cavity and do not require additional steps.[21,22]

Photogrammetry can also be helpful in the digitization of patient information. Studies have shown that the use of photogrammetric techniques can accelerate the process of loading full dental arch prostheses by integrating



Fig. 1. Graphic depiction of the photogrammetric method

A point in object space, the projective center of the optical system, and the corresponding point in image space lie on a single straight line. Thus, a spatial point can be located at the intersection of 2 rays projected from 2 different camera positions.[9] Original figure was prepared by DP.

photogrammetric 3D models of dental impressions with models obtained from cone beam computed tomography (CBCT).[23,24]

Photogrammetry, when used to measure cephalometric parameters, can accurately predict the values obtained using traditional cephalometric methods, potentially limiting their use to areas of specific interest.[25–27] It also provides a viable approach for objectively assessing changes in facial appearance following dental reconstruction.[28]

There has been a noticeable increase in the number of articles describing the use of photogrammetry in dentistry and maxillofacial surgery in recent years. Authors tend to focus on its application in prosthetic dentistry and anthropological measurements.[16–18]

Since the last systematic review assessing the accuracy and precision of photogrammetry in dentistry was published, many articles on this subject have emerged. Hence, a decision was made to conduct a systematic review that evaluates the use of photogrammetry in dentistry with the aim of improving the understanding and validation of its relevance in dental medicine. This review was structured based on the following PICO question: "How accurate and reliable are photogrammetric techniques, either used alone or compared with traditional methods, in dentistry?"

Material and methods

Eligibility criteria

Based on the Preferred Reporting Items for Systematic reviews and Meta-Analyses (PRISMA) guidelines,[29] a description of the criteria used to assess the eligibility of each study for inclusion in our systematic review was prepared. To be included, an article had to:

- be an original study published in a scientific, peer-reviewed journal;
- be published in English;
- describe the type of dental, prosthodontic or cephalometric measurements performed using a photogrammetric technique;
- describe the accuracy or reproducibility of the above-mentioned measurements.

This systematic review did not include articles for which full study data was unavailable, such as conference abstracts, or articles of uncertain publication status, for example, non-peer-reviewed preprints. Case studies and technical notes were also excluded.

Information sources

At the beginning of December 2023, 4 databases were searched: Embase; Google Scholar; PubMed®; and Web of Science. The first 2,000 articles sorted by relevance in Google Scholar were scoped, as this number has been shown to be sufficient for a systematic review.[30] Although the cited study refers to only the first 200 results, the decision was made to extend the search in order to include studies that might not have been indexed in other databases, considering the novelty of the topic.

Due to the over 12-month time period between the primary search and manuscript submission, additional searches were conducted at the end of February 2025 to incorporate newly reported findings. Re-running searches before manuscript submission is necessary, according to the Cochrane Handbook for Systematic Reviews of Interventions – C37, to identify and incorporate the latest studies into the systematic review.[31]

Search strategy

During the database search, the following queries were used:

- PubMed®:
  - ("photogrammetry"[MeSH Terms] OR "photogrammetry"[All Fields]) AND ("dentistry"[MeSH Terms] OR "dentistry"[All Fields] OR "dentistry s"[All Fields]);
  - ("photogrammetry"[MeSH Terms] OR "photogrammetry"[All Fields]) AND "dental implants"[All Fields];
  - ("photogrammetry"[MeSH Terms] OR "photogrammetry"[All Fields] OR "stereophotogrammetry"[All Fields]) AND ("dentistry"[MeSH Terms] OR "dentistry"[All Fields] OR "dentistry s"[All Fields]);
  - ("photogrammetry"[MeSH Terms] OR "photogrammetry"[All Fields] OR "stereophotogrammetry"[All Fields]) AND "dental implants"[All Fields].
- Embase:
  - ('photogrammetry'/exp OR 'photogrammetry') AND ('dentistry'/exp OR 'dentistry');
  - ('photogrammetry'/exp OR 'photogrammetry') AND ('dental implants'/exp OR 'dental implants');
  - ('stereophotogrammetry'/exp OR 'stereophotogrammetry') AND ('dentistry'/exp OR 'dentistry');
  - ('stereophotogrammetry'/exp OR 'stereophotogrammetry') AND ('dental implants'/exp OR 'dental implants').
- Web of Science:
  - Photogrammetry AND Dentistry (All Fields);
  - Photogrammetry AND Tooth implants (All Fields);
  - Stereophotogrammetry AND Dentistry (All Fields);
  - Stereophotogrammetry AND Tooth implants (All Fields);
- Google Scholar:
  - Photogrammetry AND Dentistry;
  - Photogrammetry AND Tooth implants;
  - Stereophotogrammetry AND Dentistry;
  - Stereophotogrammetry AND Tooth implants.

Searches were not limited by language. We did not use any filters during the database search. For the re-run searches, a publication date filter was applied to include studies published since January 2024.

Considering the novelty of this review's subject and PICO question, "photogrammetry" and "stereophotogrammetry" were used as keywords that would ensure the identification of articles regarding this specific topic, and "dentistry" was used as a broad MeSH and Emtree term allowing the identification of articles within this entire specialty. The keyword "dental implants" was introduced to identify studies regarding dental implant measurements. Although it is included in the "dentistry" MeSH tree, it is not included in the corresponding Emtree hierarchy; therefore, its use was necessary. This search strategy, along with the inclusion of the first 2,000 results from Google Scholar, although more time-consuming, was intended to increase sensitivity by using very broad terms.

Selection process

Four authors (OWP, JG, KM, WT) were responsible for title screening. Each of the 4 authors searched an assigned database: Google Scholar (OWP); Web of Science (JG); Embase (KM); and PubMed® (WT). All authors performed their searches independently. The articles with titles related to the use of photogrammetry in dentistry as a measurement method were collected. Duplicates among the retrieved articles were then removed using Mendeley software (Elsevier, London, UK). Articles retrieved by the other authors were reviewed by OWP, and articles retrieved by OWP were reviewed by JG.

Four authors (OWP, KM, WT, JG) were responsible for screening the abstracts of articles identified in the first step. Each author independently reviewed an equal number of articles, which were assigned in alphabetical order. Using Mendeley software, articles were marked with the tags "included" if accepted and "excluded" if rejected. After the initial evaluation of the abstracts was completed, OWP reviewed the articles assessed by the other authors, and articles assessed by OWP were reviewed by JG.

Data collection

Three authors (KM, WT, JG) were responsible for collecting data from the initially included articles. For this step, we used an Excel spreadsheet in which each paper was evaluated according to the eligibility criteria. Each author collected data independently from an equal number of articles assigned in alphabetical order. The initially collected data was then reviewed by OWP.

Data items

The following data was extracted from each study: population, understood as the group of patients or the type of artificial model used in the study; type of measurements – cephalometric measurement, implant position measurement or dental arch measurement; method of measurement, understood as the technique used to generate the models from which measurements were taken or the technique used to perform direct measurements; and the software used to process photogrammetric data. The accuracy and reproducibility of the measurements were considered the outcomes of the reviewed studies.

Assessment of the risk of bias

Three authors (KM, WT, DP) were responsible for the risk-of-bias assessment of the studies. Each author independently assessed an equal number of studies, which were assigned alphabetically. Each data source was evaluated for risk of bias using the original Quality Assessment of Diagnostic Accuracy Studies (QUADAS-1) tool, which consists of 14 items (Table 1). QUADAS-1 is a validated, evidence-based quality assessment tool.[32,33] Each item was rated as "+", indicating a positive response, or "–", indicating a negative response. The initial assessment was reviewed by OP, and any disagreements were resolved through discussion between the authors.

Effect measures

For the accuracy and reproducibility outcomes, the extracted data showed considerable heterogeneity in the types of effect measures reported. Therefore, all effect measures related to the reproducibility or accuracy of photogrammetric measurements were taken into consideration. Effect measures for reproducibility outcomes included the intraclass correlation coefficient (*ICC*), Pearson's correlation coefficient, root mean square (RMS), mean deviation, and interquartile range. For accuracy, the effect measures included RMS deviation, percentage of measurements within a specified error range, mean deviation, and standard deviation of distances.

Data synthesis

The collected data from each article was then assigned to one of the 3 categories depending on the type of measurement performed: cephalometric measurement (n = 12), dental implant position measurement (n = 18), or dental arch measurement (n = 14), and was further classified according to whether accuracy or reproducibility was assessed.

The data was then sorted and divided into 6 independent groups to properly visualize and describe the synthesized information: cephalometric studies that measured reproducibility; cephalometric studies that measured accuracy; dental implant studies that measured reproducibility; dental implant studies that measured accuracy; dental arch studies that measured reproducibility; and dental arch studies that measured accuracy. A narrative synthesis was performed for each group. Due to the considerable variation in the types of measured effects, a statistical analysis of heterogeneity was not conducted.

Table 1. Items included in the Quality Assessment of Diagnostic Accuracy Studies (QUADAS-1) tool used for the assessment of the risk of bias of studies included in the review

Item	Question
D1	Was the spectrum of patients representative of the patients who will receive the test in practice?
D2	Were selection criteria clearly described?
D3	Is the reference standard likely to correctly classify the target condition?
D4	Is the time period between reference standard and index test short enough to be reasonably sure that the target condition did not change between the two tests?
D5	Did the whole sample or a random selection of the sample, receive verification using a reference standard?
D6	Did patients receive the same reference standard regardless of the index test result?
D7	Was the reference standard independent of the index test (i.e., the index test did not form part of the reference standard)?
D8	Was the execution of the index test described in sufficient detail to permit replication of the test?
D9	Was the execution of the reference standard described in sufficient detail to permit its replication?
D10	Were the index test results interpreted without knowledge of the results of the reference standard?
D11	Were the reference standard results interpreted without knowledge of the results of the index test?
D12	Were the same clinical data available when test results were interpreted as would be available when the test is used in practice?
D13	Were uninterpretable/intermediate test results reported?
D14	Were withdrawals from the study explained?

Results

During the database searches, 4,914 records were initially identified. After title screening for studies describing the use of photogrammetry in dentistry and the removal of duplicates, 490 articles remained. Abstract screening was then conducted, during which 396 papers were excluded, leaving 94 articles. Subsequently, the full texts of the remaining articles were evaluated according to the eligibility criteria. During this phase, 50 articles were excluded, and 44 articles were included in the final review (Fig. 2). Despite meeting the eligibility criteria, 1 study was excluded because of its insufficient sample size – a sample of 6 men and 6 women – which resulted in statistically non-significant findings with high *p*-values for all but one outcome.[34] Therefore, this study, despite meeting the eligibility criteria, did not provide sufficiently informative data regarding the accuracy or reliability of photogrammetric measurements and was excluded, in accordance with the PRISMA explanation and elaboration paper.[29]



Fig. 2. Study flow diagram based on the Preferred Reporting Items for Systematic reviews and Meta-Analyses (PRISMA) 2020 flow diagram for new systematic reviews that include searches of databases and registers only

Risk of bias

For each study included in this review, the assessment of the risk of bias was conducted and presented graphically, with "+" indicating a positive response and "–"indicating a negative response (Table 2). Most studies received positive ratings for items related to the reference standard and the execution of the index and reference tests (D3–D9), with low-risk ratings ranging from 79.5% for D7 to 97.7% for D4. In contrast, high-risk ratings were more frequent in domains D1 (79.5%), D2 (68.2%), D10 (68.2%), D11 (54.5%), and D12 (52.3%). The highest proportions of high-risk ratings were observed in D13 (95.5%) and D14 (97.7%).

Reproducibility of cephalometric measurements

Seven studies described the reproducibility of photogrammetric measurements in the context of cephalometric measurements (Table 3). All studies described this technique as reproducible, with the exception of a single measurement reported in 1 study – the distance from the incisal edge to the distal canine.[25,28,35–39]

Table 2. Risk-of-bias assessment for each study included in the systematic review

Study	Risk of bias													
	D1	D2	D3	D4	D5	D6	D7	D8	D9	D10	D11	D12	D13	D14
Al-Khatib et al. 2011[64]	–	+	+	+	+	+	–	–	+	+	–	+	–	–
Al-Khatib et al. 2012[63]	+	+	+	+	+	+	+	+	+	+	+	+	–	–
Arapović-Savić et al. 2019[65]	–	–	+	+	+	+	+	+	+	+	+	+	–	–
Artopoulos et al. 2014[40]	–	–	+	+	+	+	+	+	+	–	+	–	–	–
Bergin et al. 2013[60]	–	–	+	+	+	+	+	+	+	+	+	+	–	–
Bratos et al. 2018[45]	–	–	+	+	+	+	+	+	+	–	–	–	–	–
Brons et al. 2013[35]	+	+	+	+	+	+	+	+	+	–	–	+	–	+
Castillo et al. 2018[25]	+	–	+	+	+	+	–	+	–	–	+	+	–	–
Ceinos et al. 2016[36]	–	+	–	+	–	–	–	–	–	–	–	+	–	–
Chaudhary et al. 2024[75]	+	+	+	+	+	+	+	+	–	–	–	–	–	–
Cheng et al. 2024[52]	–	–	+	+	+	+	+	+	+	–	–	–	–	–
Codari et al. 2016[70]	–	–	+	+	+	+	+	+	+	+	+	+	–	–
Di Blasio et al. 2025[43]	–	–	–	+	+	+	+	+	–	–	–	+	–	–
Estibalez-Recasens et al. 2025[56]	+	+	+	+	+	–	–	+	+	+	–	+	–	–
Faria et al. 2025[55]	–	+	+	+	+	+	+	+	+	–	–	–	–	–
Fu et al. 2017[66]	+	+	+	+	+	+	+	+	+	–	+	+	–	–
Goellner et al. 2010[67]	–	–	+	+	+	+	+	+	+	–	–	+	–	–
Hartmann et al. 2025[39]	–	+	+	+	+	+	+	+	+	–	–	–	+	–
Kamel et al. 2022[71]	–	–	+	+	+	–	+	+	+	+	+	–	–	–
Keil et al. 2025[38]	–	+	+	+	+	+	+	+	+	–	–	–	+	–
Lincoln et al. 2016[41]	–	–	+	+	+	+	+	–	+	+	+	–	–	–
Liu et al. 2024[54]	–	–	+	+	+	+	+	+	+	–	–	–	–	–
Ma et al. 2021[46]	–	–	+	+	+	+	+	+	+	+	+	+	–	–
Mahmood et al. 2023[69]	–	–	+	+	+	+	+	+	–	–	+	–	–	–
Mai and Lee 2022[72]	–	+	+	+	+	+	+	+	+	–	–	–	–	–
Örtorp et al. 2005[61]	–	–	+	+	+	+	+	+	+	–	–	–	–	–
Pinto et al. 2024[47]	–	–	+	+	+	+	+	+	+	+	+	+	–	–
Pozzi et al. 2023[53]	–	–	+	+	+	+	+	+	–	+	+	+	–	–

Study	Risk of bias													
	D1	D2	D3	D4	D5	D6	D7	D8	D9	D10	D11	D12	D13	D14
Rangel et al. 2008[73]	–	–	–	+	–	–	+	+	–	–	–	–	–	–
Revilla-León et al. 2021[57]	–	–	+	+	+	+	+	+	+	–	–	–	–	–
Revilla-León et al. 2023[62]	–	–	+	+	+	+	+	+	+	+	+	+	–	–
Revilla-León et al. 2024[48]	–	–	+	+	+	+	–	+	+	–	+	–	–	–
Revilla-León et al. 2025[58]	–	–	+	+	+	+	–	+	+	+	+	–	–	–
Rivara et al. 2016[49]	+	–	+	+	+	+	+	+	+	–	–	+	–	–
Seifert et al. 2024[44]	–	+	+	+	+	+	+	+	+	–	–	+	–	–
Sterenborg et al. 2016[28]	+	+	+	+	+	+	–	+	+	–	+	+	–	–
Sterenborg et al. 2018[37]	+	+	–	–	–	–	–	+	–	–	–	+	–	–
Tohme et al. 2021[51]	–	–	+	+	+	+	+	+	+	+	+	+	–	–
Tohme et al. 2023[50]	–	–	+	+	+	+	+	+	+	–	–	–	–	–
Stuani et al. 2019[17]	–	–	+	+	+	+	–	+	+	–	–	–	–	–
Yang et al. 2023[74]	–	–	+	+	+	+	+	+	+	–	+	–	–	–
Zhou et al. 2024[59]	–	–	+	+	+	+	+	+	+	–	–	–	–	–
Zotti et al. 2022[18]	–	–	+	+	+	+	+	+	+	–	+	–	–	–
Majid et al. 2006[42]	–	–	+	+	+	+	+	+	+	–	–	–	–	–

Table 3. Reproducibility of photogrammetric cephalometric measurements

Study	Population/materials	Technique and software used for measurements	Reproducibility of measurements
Sterenborg et al.[28]	adults	photogrammetry, 3dMDface	Most measurements had a Pearson's correlation coefficient of >0.8. Measurements of the stomion–gnathion and subnasale–gnathion showed lower correlation coefficients, with the lowest values being 0.46 and 0.76, respectively.
Brons et al.[35]	pediatric patients	photogrammetry, 3dMDface	95% of the distances were 0.80 mm or less for intraobserver measurements and below 1.10 mm for interobserver measurements.
Ceinos et al.[36]	adults	photogrammetry, DermaPix	The *ICC* values for intra- and interexaminer facial measurements ranged from 0.732 to 0.976, except for measurements involving the inner part of the dental corridor, for which *ICC* values ranged from 0.598 to 0.914.
Sterenborg et al.[37]	adults	photogrammetry, 3dMDface	All measurements had a Pearson's correlation coefficient of >0.85.
Castillo et al.[25]	adults	photogrammetry, Vectra 3D	The inter- and intraexaminer Pearson's correlation coeffcient for photogrammetry averaged 0.94 and 0.92, respectively.
Hartmann et al.[39]	adults	photogrammetry, Vectra M5, 3D Scanner App	Photogrammetric models generated using smartphones showed higher interobserver reliability for overall landmark-to-landmark deviation, superior alignment and higher volumetric accuracy compared with models generated using the TrueDepth camera.
Keil et al.[38]	mannequin	photogrammetry, Heges 3D, Agisoft Metashape	The lowest precision values for photogrammetry were achieved at 500 lx. Photogrammetry showed the highest precision at 5,000 lx and slightly lower precision at 20,000 lx.

ICC – intraclass correlation coefficient.

The highest Pearson's correlation coefficient was reported in a study evaluating cephalometric measurements in relation to dental measurements. The mean value was 0.94 for interexaminer reliability and 0.92 for intraexaminer reliability, indicating excellent reliability of all measurements made.[25]

Two studies assessed changes in the vertical dimension of occlusion and showed that the photogrammetric technique achieved correlation coefficients above 0.80 for all measurements, indicating good reliability.[28,37]

One study of cephalometric measurements described photogrammetry as having lower precision. Although facial measurements showed inter- and intraoperator *ICC*s ranging from 0.732 to 0.976, indicating good to excellent reliability, the *ICC* values for measurements of the dental corridor ranged from 0.598 to 0.914, indicating lower repeatability. The lowest *ICC* reported for measurements of the dental corridor was observed for the incisal edge – right premolar segment and was 0.732, with a 95% confidence interval (95% *CI*) ranging from 0.285 to 0.915.[36]

Ambient light has a significant influence on the reliability of photogrammetry-based facial models. Keil et al. showed that the photogrammetric method achieved the highest precision at 5,000 lx, slightly lower precision at 20,000 lx, and the lowest precision at 500 lx.[38]

For measurements in a pediatric population, with and without orofacial clefts, 95% of the distances were 0.80 mm or less for intraobserver measurements and below 1.10 mm for interobserver measurements. Statistical analysis showed that the age of the infants and the presence of orofacial clefts did not significantly affect the measurements, indicating that photogrammetry can be a reliable method for soft tissue assessment in infants with and without orofacial clefts.[35]

The 3D Scanner App (v. 2.1.2; Laan Consulting Corp., New York, USA) was also shown to be a reliable method for generating facial 3D models. It demonstrated good to excellent *ICC* values for linear and volumetric measurements. Additionally, photogrammetry-based models showed higher *ICC* values than models generated using TrueDepth camera of an iPhone 14 Pro (Apple Inc., Cupertino, USA).[39]

Accuracy of cephalometric measurements

Seven different studies described the accuracy of photogrammetric measurements in the context of cephalometric measurements (Table 4).[38–44]

A comparative testing study showed photogrammetry to be significantly less accurate than methods such as facial moulage and CBCT scans.[41] The study reported mean percentage differences of 1.067 for photogrammetry, 0.622 for CBCT and 0.482 for facial moulage, indicating that photogrammetry had the largest measurement error among these 3 methods.[41] The results of another comparative study indicated that even though photogrammetry was less accurate than the abovementioned methods, its accuracy was still acceptable, as more than 99% of the measured points in the photogrammetric model specimen were within the clinically acceptable limit of 1 mm compared with the CT scan.[40]

Table 4. Accuracy of photogrammetric cephalometric measurements

Study	Population/materials	Technique and software used for measurements	Accuracy of measurements
Lincoln et al.[41]	mannequins	photogrammetry, 3dMDface	Pairwise comparisons showed differences between the methods, with photogrammetry exhibiting a higher percentage error than the other 2 methods, whereas CBCT and the stone model showed similar percentage errors. Photogrammetry differed significantly from CBCT and the stone model ($p < 0.05$). The greatest differences were observed at the glabella, nasion and palpebrale inferius landmarks, whereas the data for the other points were similar to those obtained using CBCT and the stone model.
Artopoulos et al.[40]	mannequins	photogrammetry, Di3D Capture	More than 99% of the measured points for the photogrammetric specimen were within the clinically acceptable limit of 1 mm compared with the CT scan.
Majid et al.[42]	adults, mannequins	photogrammetry, software not reported	The natural features technique was found to significantly improve the accuracy of non-contact craniofacial anthropometric measurements. In the mannequin test, RMSE values were 0.8 mm and 6.0 mm for the natural features technique and control frame technique, respectively. For human faces, RMSE values were 2.0 mm and 3.0 mm, respectively.
Di Blasio et al.[43]	adults	photogrammetry, Polishape 3D	Measurements were repeatable and remained within the clinically acceptable range of ±1 mm even in the case of roll, pitch and yaw of the head, with the exception of measurements involving the trago.
Seifert et al.[44]	adults	photogrammetry, Polycam, ScandyPro, 3dMDface, EM3D	The mean surface distances for EM3D, Polycam and ScandyPro were 1.46 mm, 1.66 mm and 1.61 mm, respectively. The mean landmark distances for EM3D, Polycam and ScandyPro were 1.45 mm, 1.27 mm and 1.26 mm, respectively. ANOVA showed no significant differences among scanners for any landmark.
Hartmann et al.[39]	adults	photogrammetry, Vectra M5, 3D Scanner App	Smartphone-based approaches using both the TrueDepth camera and photogrammetry exhibited overall landmark-to-landmark distances of ≤2 mm.
Keil et al.[38]	mannequin	photogrammetry, Heges 3D, Agisoft Metashape	Single-camera photogrammetry showed the lowest RMSE values at 5,000 lx, although only random differences were observed compared with 20,000 lx.

CBCT – cone beam computed tomography; CT – computed tomography; RMSE – root mean square error; ANOVA – analysis of variance.

External factors such as ambient light and patient cooperation may influence the accuracy of photogrammetric facial model acquisition. At a light intensity of 5,000 lx, the results were most accurate, whereas at 20,000 lx they were slightly less precise, with the least accurate results observed at 500 lx.[38] Interestingly, patient head positioning affected the accuracy of photogrammetry only slightly. Regardless of head position, all measurements were within the clinically acceptable limit of 1 mm except for the trago.[43]

Two comparative studies tested different methods of acquiring photogrammetry-based facial 3D models. One showed that there were no significant differences between models generated using a smartphone-based TrueDepth camera and those generated using photogrammetry, whereas all of them exhibited overall landmark-to-landmark distances of ≤2 mm.[39] The other study analyzed discrepancies between models generated using different photogrammetry-based smartphone applications and showed no significant differences between them.[44]

Majid et al. proposed a novel photogrammetric technique based on characteristic points of the face, namely the natural features technique.[42] Its accuracy was measured using the root mean square error (RMSE) and was 0.351 mm on the x-axis, 0.381 mm on the y-axis, and 0.843 mm on the z-axis. The method was more accurate than conventional photogrammetry.[42]

Reproducibility of dental implant position measurements

Twelve different studies described the reproducibility of photogrammetric measurements in the context of dental implant positions (Table 5). All of them characterized the technique as reproducible.[45–56]

Table 5. Reproducibility of photogrammetric dental implant position measurements

Study	Population/materials	Technique and software used for measurements	Reproducibility of measurements
Pinto et al.[47]	dental casts	photogrammetry, ICam4D, PIC Dental	Mean RMS repeatability values for the model with 4 implants were 5.18 µm for PIC and 7.01 µm for ICAM. For the model with 6 implants, the corresponding values were 13.88 µm and 8.67 µm.
Tohme et al.[51]	dental casts	photogrammetry, PIC Dental	The mean angular deviation did not differ significantly among the impression techniques and among ISBs within DIOS and DPG. The mean RMS differed significantly among impression techniques ($p < 0.001$) and was lower with DPG (0.014 ±0.013). The mean RMS for 3D deviation within ISBs did not differ significantly for DIOS ($p = 0.615$) or DPG ($p = 0.666$).
Ma et al.[46]	dental casts	photogrammetry, ICam4D	Median precision values were 2.00 µm (*IQR*: 1.65), 36.00 µm (*IQR*: 9.95) and 29.40 µm (*IQR*: 4.80) for the photogrammetry, IOS and CNV groups, respectively. The PG group showed significantly higher precision than the IOS ($p < 0.001$) and CNV groups ($p < 0.001$).
Tohme et al.[50]	dental casts	photogrammetry, PIC Dental	The mean RMS values for 3D deviations of the whole scan body ($p < 0.001$) and the flat angled surface ($p < 0.001$) differed significantly among the techniques and were significantly lower with photogrammetry. Mean global angular deviations also differed significantly among the techniques and were lower with photogrammetric acquisition ($p < 0.001$).
Bratos et al.[45]	dental casts	photogrammetry, iWitnessPRO	For photogrammetry, the grand mean standard error was 26.8 µm (95% *CI*: 18.1–51.4 µm).
Revilla-León et al.[48]	dental casts	photogrammetry, Geomagic	Levene's test showed no significant differences in mean precision among the tested groups ($p = 0.197$).
Rivara et al.[49]	dental casts	photogrammetry, photomodeler, EOS System	The tests showed an average precision of 4 mm for implant positions and less than 1 degree for implant axes.
Cheng et al.[52]	dental casts	photogrammetry, ICam4D	Photogrammetry exhibited the lowest RMS 3D and angular deviations at all implant positions compared with CNV, IOS and MIOS. It also showed lower deviations at the D1 distance than CNV, IOS and MIOS.
Pozzi et al.[53]	dental casts	photogrammetry, PIC Dental	A significantly lower *SD* of the mean 3D deviations was observed with photogrammetry than with IOS (17.7 µm vs. 37.1 µm, $p < 0.0001$). The *SD* of angular measurements was also significantly lower with photogrammetry than with IOS (0.04° vs. 0.14°, $p < 0.0001$).
Liu et al.[54]	dental casts	photogrammetry, PIC Dental	Photogrammetry showed the highest precision for distance and angulation measurements. Its linear, angular and RMS deviation values (9–11 µm, 0.06–0.08°, 8–10 µm, respectively) were significantly lower than those for open tray impression and IOS according to ANOVA.
Faria et al.[55]	dental casts	photogrammetry, PIC Dental	For photogrammetry, no significant differences ($p = 0.075$) were found among the 3 implant distributions evaluated. High precision was observed for all devices analyzed, with no significant differences among them ($p = 0.097$). Overall precision, expressed as RMS, was 12.42 µm (95% *CI*: 11.33–13.50 µm) for photogrammetry.
Estibalez-Recasens et al.[56]	adults	photogrammetry, PIC Dental	For linear measurements, the *M* ±*SD* precision was 14.71 ±12.21 µm. For angular measurements, the *M* ±*SD* precision was 0.072 ±0.064°.

RMS – root mean square; ISBs – intraoral scan bodies; DIOS – digital intraoral scanning; DPG – digital photogrammetry; IOS – intraoral scanners; MIOS – intraoral scanning with splinting; CNV – conventional impression; 95% *CI* – 95% confidence interval; *SD* – standard deviation; *IQR* – interquartile range; *M* – mean.

Five studies comparing the precision of photogrammetry, intraoral scanning and conventional impression techniques for complete arch implant rehabilitation showed photogrammetry to be significantly more precise than the other methods. In the first study, the median precision was 2.00 µm.[46] Another study showed that, in terms of precision, the mean RMS values for 3D deviations of both the entire scan body and the flat angled surface were significantly smaller with photogrammetry. Comparison of the mean global angular deviations yielded similar results.[50] A previous study by the same author compared photogrammetry with IOS and found no differences in angular deviations or the mean RMS values of 3D deviations between the 2 techniques.[51] The most recent studies further demonstrated the superior precision of photogrammetry compared with IOS and conventional impressions.[52–54]

No significant differences in terms of precision were found in a study investigating the impact of changes in scanning distance on the accuracy and reproducibility of photogrammetry.[48]

In a study involving a coordinate-measuring machine (CMM), it was demonstrated that the precision of photogrammetry was similar to that of the CMM.[45] For the groups that used the photogrammetry method, the grand mean for linear measurements was 26.8 µm; however, 2 groups – AvgR-MaxO and MaxR-MaxO – were not included in the calculation due to the limitation of the typodont in simulating living tissues. The precision of angular measurements also showed no significant differences between the groups, with grand means of 0.129° for the x angle and 0.221° for the y angle.[45]

Pinto et al. evaluated the repeatability of various digital acquisition methods for capturing implant positions in prostheses with 4 and 6 implants.[47] The study findings suggested that photogrammetry demonstrated superior repeatability compared with IOS for recording implant locations in complete arch prostheses supported by both 4 and 6 implants, but were less repeatable than laboratory scanners.[47] The most recent findings confirmed that the number of implants did not influence the precision of photogrammetry and that its precision remained very high, with an RMS of 12.42 µm.[55] Additionally, a study by Estibalez-Recasens et al. showed that the in vivo precision of photogrammetry in patients with 6 implants was similarly high.[56] The linear and angular measurements were 14.71 ±12.21 µm and 0.072 ±0.064°, respectively.[56]

Accuracy of dental implant position measurements

Fourteen studies described the accuracy of photogrammetric measurements of dental implant positions (Table 6).[45,46,48,50–54,57–62]

In 5 comparative studies, photogrammetry was shown to have high accuracy.[46,51–54] All of them showed photogrammetry to be more accurate in terms of 3D and angular discrepancies than IOS.[51,53] Additionally, three of them reported higher trueness for photogrammetry than for conventional impressions.[46,52,54]

Two studies compared photogrammetry with conventional impression techniques. In the study by Bergin et al., the mean linear measurement error was 26.2 µm for the conventional impression technique and 28.8 µm for photogrammetry, with no statistically significant difference between the methods.[60] In contrast, Zhou et al. reported greater linear and angular discrepancies for photogrammetry, with mean values of 145 ±196 µm and 0.82 ±0.88°, respectively, which were significantly greater than those obtained with the conventional impression technique.[59]

Interestingly, the study by Örtorp et al. showed that cylinder center-point distortion was smaller with the photogrammetric technique than with the conventional impression technique, whereas angular cylinder distortion was greater with photogrammetry.[61]

Four articles regarding dental implant position measurements described photogrammetry as having lower overall accuracy compared with a CMM. Revilla-León et al. measured the average 3D discrepancy of photogrammetry at 77.6 µm.[57] Their later study confirmed these results, showing a statistically significant difference between the CMM method and photogrammetry, with a trueness difference of 1.8 µm between the systems.[62] Similarly, Bratos et al. conducted a study that showed a mean linear accuracy of 55.2 µm, which was lower than that of the CMM.[45] Additionally, Bergin et al. found the CMM to have higher accuracy than photogrammetry, but the difference was not statistically significant.[60]

Two articles by Revilla-León et al. investigated how scanning distance and the reference point affect the performance of photogrammetry.[48,58] In one study, the authors compared accuracy at 3 different distances.[48] The 30-cm scanning group yielded the most accurate linear measurements, with a linear discrepancy of 17 ±4 µm and an angular discrepancy of 0.17 ±0.02°. Measurements performed at shorter (20 cm) or longer (35 cm) distances were less accurate.[48] In a study investigating the effect of the reference points, it was demonstrated that the selected reference point could also influence the accuracy of photogrammetry.[58] Three groups were created based on the implant used as the reference point: the right first molar, left canine and left first molar. The study revealed differences in linear and angular trueness between the groups, with the first right molar used as a reference point showing the highest trueness values. Despite the differences between the groups, the measured angular and linear discrepancies were 0.01 ±0.01° and 6 ±3 µm, respectively. Therefore, the impact of the observed discrepancies was unlikely to be clinically significant.[58]

Table 6. Accuracy of photogrammetric dental implant position measurements

Study	Population/materials	Technique and software used for measurements	Accuracy of measurements
Tohme et al.[51]	dental casts	photogrammetry, PIC Dental	Mean angular discrepancies differed significantly among the impression techniques ($p < 0.001$). The lowest value was observed with DPG (0.724 ±0.064°), whereas an elevated value was observed with DIOS (1.744 ±0.175°). The mean RMS 3D deviation differed significantly among the impression techniques ($p < 0.001$) and was lower with DPG (0.078 ±0.001) than with DIOS (0.536 ±0.063).
Pozzi et al.[53]	dental casts	photogrammetry, PIC Dental	For 3D deviations, IOS showed a higher mean 3D deviation (ΔEUC) than photogrammetry (52.8 µm vs. 33.4 µm, $p < 0.0001$). For angular deviations, IOS showed a slightly higher mean deviation (ΔANGLE) than photogrammetry (0.28° vs. 0.24°, $p = 0.0022$).
Ma et al.[46]	dental casts	photogrammetry, ICam4D	The median of trueness was 24.45 (*IQR*: 0.73), 43.45 (*IQR*: 6.17) and 28.70 (*IQR*: 7.90) µm for the photogrammetry, IOS and CNV groups, respectively. The photogrammetry group showed significantly higher trueness than the IOS ($p < 0.001$) and CNV groups ($p = 0.033$).
Bergin et al.[60]	dental casts	photogrammetry, iWitnessPRO	The average measurement error was 26.2 µm (95% *CI*: 15.9–36.6) for the conventional method and 28.8 µm (95% *CI*: 24.8–32.9) for photogrammetry. Overall measurement accuracy did not differ significantly between the conventional and photogrammetric methods.
Tohme et al.[50]	dental casts	photogrammetry, PIC Dental	The 3D deviation of the whole scan body was significantly lower with photogrammetry and higher with digital scanning. The mean RMS for 3D deviations of the flat angled surface was significantly lower with digital scanning and higher with photogrammetry ($p < 0.001$). The mean global angular distortion was lower with photogrammetry and higher with digital scanning ($p < 0.001$).
Revilla-León et al.[57]	dental casts	photogrammetry, ICam4D	Photogrammetry showed significantly greater y-axis distortion and a larger 3D discrepancy than all CNV and IOS groups ($p = 0.004$). The y-axis discrepancy for photogrammetry was 73.7 µm, and the 3D discrepancy was 77.6 µm.
Bratos et al.[45]	dental casts	photogrammetry, iWitnessPRO	The grand mean of accuracy of linear measurements for photogrammetry was 55.2 µm (95% *CI*: 8.8–130.8 µm). Angular measurements showed a similar patter to linear measurements.
Revilla-León et al.[48]	dental casts	photogrammetry, Geomagic	At a scanning distance of 20 cm, the linear discrepancy was 25 ±4 µm and the angular discrepancy was 0.15 ±0.01°. At 30 cm, the corresponding discrepancies were 17 ±4 µm and 0.17 ±0.02°. At 35 cm, they were 27 ±6 µm and 0.19 ±0.01°, respectively.
Revilla-León et al.[58]	dental casts	photogrammetry, PIC Dental	A linear discrepancy of 6 ±3 µm and an angular discrepancy of 0.01 ±0.01° were reported.
Örtorp et al.[61]	dental casts	photogrammetry, Zeiss Prismo VAST	Photogrammetry reproduced the x-axis and 3D parameters more accurately than the plaster technique when cylinder centre-point distortion was assessed. Angular cylinder distortion was greater with photogrammetry than with the impression technique.
Revilla-León et al.[62]	dental casts	photogrammetry, PIC Dental	Photogrammetry showed an overall scanning accuracy value of 20.15 ±25.41 µm.
Cheng et al.[52]	dental casts	photogrammetry, ICam4D	Photogrammetry showed lower RMS 3D discrepancies at all implant positions, reduced RMS 3D bias at implant positions #7 and #3, and lower angular bias at implant positions #7, #5 and #3. However, photogrammetry showed larger linear errors for the D3 and D4 distances.
Zhou et al.[59]	dental casts	photogrammetry, ICam4D	The overall *M* ±*SD* distance deviation for photogrammetry was 149 ±196 µm, and the angular deviation was 0.82 ±0.88°. Using 150 µm as the clinically acceptable error range, 69.7% of distance deviations were below 150 µm.
Liu et al.[54]	dental casts	photogrammetry, PIC Dental	The linear, angular and RMS trueness values for photogrammetry (31–39 µm, 0.22–0.28°, 20–23 µm) were significantly lower ($p < 0.001$) than those for open tray impression and IOS.

Reproducibility of dental arch dimension measurements

The reproducibility of dental arch dimension measurements was reported in 7 studies (Table 7).[17,18,63–67] Two of them reported *ICC* values above 0.90 for all measurements performed,[18,65] indicating excellent reliability.[68] Four studies indicated good agreement or values ranging from good to excellent, depending on the measurement, with *ICC* values varying from 0.71 to 0.998.[63,64,66,67] One study reported lower reproducibility, with a repeatability coefficient of 0.474.[17]

One of the studies measured validity and reliability using crown widths and selected dental arch dimensions.[63] The *ICC* values varied between both measurements – the mean *ICC* was 0.88 for crown widths, whereas the mean *ICC* for dental arch dimensions was 0.97.[63] A similar range of *ICC* values (0.879–0.998) was reported in a study that focused on measuring the mesiodistal crown diameter, arch width and arch perimeter.[66]

Table 7. Reproducibility of photogrammetric dental arch dimension measurements

Study	Population/materials	Technique and software used for measurements	Reproducibility of measurements
Arapović-Savić et al.[65]	dental casts	photogrammetry, OP4D	Correlation coefficients were close to 1 for all measurements.
Zotti et al.[18]	dental casts	photogrammetry, 3DF Zephyr Free	*ICC* was 0.99 for each pair of measurements, indicating almost perfect agreement between the 2 measurements.
Goellner et al.[67]	adults	photogrammetry, ARAMIS	*ICC*s ranged from "good" (0.71 at 3 N) to "optimal" (0.98 at 18 N).
Stuani et al.[17]	dental casts	photogrammetry, 3DF Zephyr Free	The precision between the digital model measurements showed a repeatability coefficient of 0.474.
Fu et al.[66]	dental casts	photogrammetry, MeshLab	The intrarater *ICC*s ranged from 0.879 to 0.998 for the digital measurements, indicating good intrarater reliability.
Al-Khatib et al.[64]	adults	photogrammetry, software not reported	Precision was generally high, with all *ICC* values >0.80.
Al-Khatib et al.[63]	dental casts	photogrammetry, Australis software	The mean *ICC* was 0.88 for crown widths and 0.97 for dental arch dimensions.

The widest range of *ICC* values was observed in a study focused on evaluating tooth displacement under a low horizontal tensile force (≤18 N), with the lowest *ICC* value of 0.71 at a force of 3 N and the highest value of 0.98 at a force of 18 N.[67]

Good reliability, with all *ICC* values >0.80, was reported in the study focused on investigating tooth size and dental arch dimensions in Malays using a stereophotogrammetric system.[64]

Two studies compared photogrammetric measurements obtained using 3D scanning systems and confirmed their high reproducibility.[18,65] The first study reported *ICC* values at 0.99 for each pair of measurements obtained with photogrammetry, indicating almost perfect agreement between the 2 measurements.[18] The second study noted similarly high *ICC* values for different measurements, including mesiodistal widths of individual teeth, mesiodistal widths of two-teeth segments, and intercanine and intermolar distances in both the maxilla and mandible, with correlations close to 1 in all cases and the lowest *ICC* value being 0.9737.[65] These studies indicated excellent reproducibility of photogrammetry.

In contrast to all the abovementioned studies, 1 study reported low reproducibility of photogrammetric measurements, with a repeatability coefficient of 0.474.[17]

Accuracy of dental arch dimension measurements

Thirteen studies evaluated the accuracy of photogrammetry in comparison with various measurement methods (Table 8).[17,18,63–66,69–75]

Three studies found no statistically significant differences between the results when photogrammetry was compared with other reliable measurement techniques.[64,66,69] When photogrammetric measurements were compared with those performed on dental casts using calipers, the statistical analysis showed no significant differences between the 2 methods except for the lower arch perimeter. Additionally, photogrammetric measurements showed clinically acceptable accuracy, with differences between measurements of less than 0.5 mm.[64,66,69] This was further confirmed by 3 other studies that investigated the accuracy of the photogrammetric technique and demonstrated clinically acceptable accuracy, with errors below 0.5 mm.[18,65,74]

A study by Kamel et al. compared the accuracy of photogrammetry and an extraoral scanner in assessing the marginal fit of crowns.[71] The findings showed no statistically significant difference between the 2 methods, and the accuracy of photogrammetry was deemed to be clinically acceptable, with discrepancies below 120 μm.[71]

Two studies that compared photogrammetric measurements with measurements performed on plaster impressions reported differences between the methods.[63,73] One study observed differences in measurements involving the second premolars, canines and arch perimeter when comparing values obtained using calipers and photogrammetry, with photogrammetry showing larger discrepancies. However, all measurements remained within a clinically acceptable error range of less than 0.5 mm.[63] The most recent study by Chaudhary et al. showed no statistically significant differences between measurements made on gypsum models and photogrammetric digital models for intercanine distances and distances from the first premoral to the first molar in both quadrants.[75] Additionally, there was no significant difference in the volumetric comparison.[75]

One study assessed photogrammetry against measurements obtained using CBCT.[70] No statistically significant differences were noted between the results obtained using 3D dentofacial imaging and the CBCT reference, except for a single distance.[70]

An analysis of different approaches to facial scanning concluded that the use of extraoral markers significantly enhanced the accuracy of 3D dentofacial images obtained using photogrammetry.[72]

Table 8. Accuracy of photogrammetric dental arch dimension measurements

Study	Population/materials	Technique and software used for measurements	Accuracy of measurements
Arapović-Savić et al.[65]	dental casts	photogrammetry, OP4D	For the mesiodistal widths of individual teeth, the mean error was <0.2 mm. For the mesiodistal widths of two-tooth segments, errors were approx. 0.2 mm. For intercanine and intermolar distances in both the maxilla and mandible, errors were <0.3 mm, with a mean value of 0.2741 mm. The recorded differences were considered clinically negligible.
Yang et al.[74]	adults, dental casts	photogrammetry, PhotoCatch	The majority of reconstructed points fell within the ±0.2 mm error range. The reported values were 80.9 µm ±3.1% for an f/2.5 lens with a polarizer and 82.4 µm ±2.7% for an f/5.6 lens with a polarizer.
Mahmood et al.[69]	dental casts	photogrammetry, Agisoft Metashape	The mean differences between measurements ranged from –0.042 mm to 0.355 mm, were clinically acceptable (<0.5 mm) and were not statistically significant ($p > 0.05$).
Zotti et al.[18]	dental casts	photogrammetry, 3DF Zephyr Free	For the various superimpositions, the *M* ±*SD* differences between the 2 meshes were –0.02 ±0.077 mm for photogrammetry vs. dental extraoral laboratory scanner and 0.009 ±0.087 mm for photogrammetry vs. IOS, indicating good agreement among models obtained using the different acquisition methods.
Mai and Lee[72]	adults	photogrammetry, Di3D Capture	The 3D positional discrepancy was 0.959 mm (*SD*: 0.043) for tooth-based matching and 0.078 mm (*SD*: 0.041) for marker-based matching. Occlusal angle deviation was 3.187° (*SD*: 0.147) for tooth-based matching and 0.955° (*SD*: 0.350) for marker-based matching. Dental midline deviation was 0.393 mm (*SD*: 0.257) for tooth-based matching and 0.054 mm (*SD*: 0.074) for marker-based matching.
Rangel et al.[73]	adults	photogrammetry, 3dMDface	The average distance between the matched datasets was 0.35 mm (*SD*: 0.32 mm), indicating acceptable agreement between the datasets.
Kamel et al.[71]	dental casts	photogrammetry, 3DF Zephyr Free	The mean marginal gap obtained using photogrammetric scanning was 99.6 ±28.3 µm.
Stuani et al.[17]	dental casts	photogrammetry, 3DF Zephyr Free	The accuracy of digital model measurements was characterized by an *SD* of ±0.171 mm.
Fu et al.[66]	dental casts	photogrammetry, MeshLab	The mean differences between measurements ranged from 0.011 mm to 0.402 mm. The mean differences between measurements obtained by photogrammetry and on dental casts were not significant, except for the lower arch perimeter. All differences were regarded as clinically acceptable (<0.5 mm).
Al-Khatib et al.[64]	adults	photogrammetry, software not reported	Most of the mean differences between calliper and photogrammetric measurements were not statistically significant ($p > 0.05$). The differences ranged from 0.02 mm to 0.2 mm.
Codari et al.[70]	adults	photogrammetry, Vectra 3D	The median distance between ROI point clouds was 0.59 mm (0.43–0.73 mm). The median percentage of corresponding points was 86.1% (81.6–92.1%).
Al-Khatib et al.[63]	dental casts	photogrammetry, Australis software	A high degree of validity was observed, with a mean difference of <0.5 mm between direct and indirect measurements.
Chaudhary et al.[75]	dental casts	photogrammetry, 3DF Zephyr Free	Statistical analysis showed no significant differences between photogrammetric and gypsum models in the C–C distance, the distance between the mesial marginal ridge of the first premolar and the distal marginal ridge of the first molar in the first and second quadrants, or volumetric measurements.

ROI – region of interest.

Discussion

The findings suggest that photogrammetry may be a promising measurement method in dentistry for assessing cephalometric measurements, dental implant positions and dental arch dimensions. However, further studies are needed to provide robust evidence of its clinical utility before its routine implementation in dental practice can be recommended.

In the case of cephalometric measurements, it should be noted that indirect measurements, such as 3D face scans, can be more precise than direct methods, such as measurements performed using calipers. Direct measurements can comprise soft tissues or trigger muscle reflexes, such as blinking, resulting in reduced accuracy.[76] Photogrammetric reconstructions, being non-contact, do not have this disadvantage.

Additionally, once a photogrammetric 3D scan of a patient's face is acquired, the generated digital model can be stored on a computer, similarly to 3D models reconstructed from CT scans.[77] Therefore, measurements can be taken from the model without the patient's presence. This feature may be particularly useful in situations requiring repeated measurements and when integration with other digital 3D models is needed (Fig. 3).

The reviewed articles reported that photogrammetry is also a reliable method for soft tissue measurements in pediatric patients, regardless of the children's age or the presence of orofacial clefts. This may make it a valuable tool for dentists, who often face difficulties when performing

uncomfortable but necessary measurements of the face, dental arches and other structures in children.

Photogrammetric measurements of dental arches are reliable, with high *ICC* values of up to 0.99 reported in most studies. However, one study showed low precision. The majority of studies indicated that the discrepancies between photogrammetric measurements and standard methods were not statistically significant and that the measurement errors were acceptable, with values below 0.5 mm.

Given the high accuracy and reproducibility of photogrammetry, plaster dental arch models can be accurately digitized using this technique. Therefore, after further research into the capabilities and limitations of photogrammetry, especially in clinical settings, this method may become a potential alternative to more expensive extraoral scanners in the future. Additionally, because of its relatively low cost, photogrammetry can be used by students to obtain 3D dental casts, as the technique primarily requires appropriate photographic acquisition skills.[78] Photogrammetry may also be useful for teaching purposes when a university cannot afford expensive scanners, enabling students to learn digital dental workflows during their training. However, the limitations of this technique should be considered; for example, direct scanning of gingiva is not possible.

Photogrammetry appears to be a highly reproducible technique for measuring dental implant positions in terms of both linear and angular measurements.[50–54] Studies indicate that photogrammetry is more accurate than IOS but less accurate than the CMM.[45,51,53,57,60,62] In the case of dental impressions, the results are inconclusive.[59–61] Additionally, it was shown that working conditions, such as the distance between the camera and the scanned object, have a significant impact on measurement accuracy.[48]



Fig. 3. Exemplary photogrammetric 3D digital model of a mannequin's face with cephalometric landmarks and distances between the glabella and nasion, the left and right masseters, and the subnasale and menton measured in cm

The three-dimensional photogrammetric digital model was created using Polycam software (Polycam Inc., San Francisco, USA) and measured using Blender software (Blender Studio, Amsterdam, the Netherlands). Original figure was prepared by OWP.

Even though photogrammetry demonstrated superior repeatability compared with IOS for capturing implant positions in prostheses, some evidence suggests that its precision, unlike that of laboratory scanners, decreases as the number of implants increases. This may indicate that photogrammetry is a practical option for cases involving fewer implants, but for more complex full-arch reconstructions, laboratory-based techniques might offer better reliability.[79] Clinicians may consider combining techniques, using photogrammetry for its non-invasive and efficient workflow while relying on laboratory scanners when greater precision across multiple implants is necessary.[24]

The higher accuracy of photogrammetry compared with IOS for recording dental implant positions may be attributed to the fact that photogrammetry is less affected by the presence of blood and saliva.[19,20] However, because the photogrammetric method used for implant position measurement relies on reference markers, the resulting 3D digital model does not capture the surrounding soft tissues. In this context, integrating both methods may be advantageous.

Integrating photogrammetry with other methods of 3D scanning, such as IOS and CBCT, can provide complementary information and enable the creation of virtual patient models; therefore, these combined techniques can be utilized in treatment planning and diagnosis. This approach may help to deliver immediate complete arch implant-supported fixed interim and definitive prostheses in suitable patient groups, thereby reducing the time required to produce these prostheses.[80–82]

The use of photogrammetry in clinical practice, particularly for implant-supported prostheses, requires careful consideration of specific conditions, especially the scanning distance. Although in a reviewed study it was confirmed that 30-cm distance provided the greatest accuracy, the optimal distance may vary depending on the environmental conditions, such as brightness and shadowing.[48,83] This is an important limitation of photogrammetry that requires further research to determine the specific conditions under which images should be taken.

In addition to its efficacy, photogrammetry is also a non-invasive, painless, comfortable, and quick method, which represents an important advantage for a diagnostic tool.[84] These characteristics are particularly attractive in pediatric dentistry, where patients are often less willing to cooperate during dental procedures.[85] To obtain a high-quality photogrammetric 3D model, the patient should remain as still as possible during the acquisition of consecutive images, as blurred pictures can decrease measurement accuracy.[86] This, in the context of working with children, may pose a challenge for the operator.

Regarding the ALARA (as low as reasonably achievable) principle, photogrammetry has an advantage over X-ray-based imaging techniques. It can be used for cephalometric measurements instead of CT scans or radiographic examinations and, potentially in the future, for measurements of dental arch dimensions as an alternative to CBCT. These findings are especially important for pregnant women and children, as they are particularly vulnerable to radiation.[72,87] However, it is important to consider that photogrammetry can capture only the surface of an object, unlike CBCT, which can be used to reconstruct structures located within the human body, such as the inferior alveolar canal.[88]

Although creating photogrammetric facial models and taking dental implant position measurements are relatively straightforward processes, digitization of dental arches requires an additional step. First, an impression must be taken, on the basis of which a physical model of the dental arches is prepared; this model can subsequently be scanned and digitized using photogrammetry. For dental implants, this technique captures only their positions, without the surrounding tissues, unlike IOS. Therefore, the digitization of the gingiva and implant positions in edentulous patients requires the combination of these methods. These are important limitations of photogrammetry, as they can potentially prolong scanning time and require more effort from the operator compared with conventional methods.

Other limitations of photogrammetry in dentistry require further research, including the establishment of standardized measurement conditions, such as appropriate lighting. From a pediatric perspective, the need for patient to remain still may be problematic. The potential reductions in precision in more complex reconstructions involving larger numbers of implants should also be considered, suggesting that photogrammetry may currently have a complementary rather than replacement role relative to laboratory scanners.

It is also important to acknowledge that this review is based mostly on ex vivo studies using artificial models. Therefore, its results cannot be directly extrapolated to clinical practice. Overall, the high accuracy and reproducibility reported for photogrammetry suggest that it may be a useful tool for performing cephalometric measurements, recording implant positions and measuring dental arches under laboratory conditions; however, its effect on clinical outcomes remains undetermined.

Conclusions

Photogrammetry has proved to be a precise and accurate method for cephalometric measurements, recording the position of dental implants and measuring dental arch dimensions. Its reliability is comparable to that of other established methods currently used in dentistry, such as IOS and traditional casts. In some cases, particularly when compared to IOS, photogrammetry showed superior outcomes. Moreover, photogrammetry may be more economically viable than other techniques.

Photogrammetry is a promising method with considerable potential in dentistry. However, it should be noted that only one of the reviewed studies was conducted in a clinical setting; therefore, photogrammetry cannot yet be considered the preferred method for the measurements discussed in this review. To assess this limitation, photogrammetric techniques should be further evaluated in clinical settings, particularly in view of the promising results obtained in laboratory conditions. We hope that this research will provide useful information and encourage clinicians to further investigate the applicability of photogrammetry in dental practice.

Ethics approval and consent to participate

Not applicable.

Data availability

Not applicable.

Consent for publication

Not applicable.

Use of AI and AI-assisted technologies

Not applicable.

ORCID iDs

Oliwier Wojciech Pioterek https://orcid.org/0009-0000-7896-9266
Katarzyna Majka https://orcid.org/0009-0002-7497-4265
Jakub Goździewicz https://orcid.org/0009-0003-4507-3043
Wiktoria Taurogińska https://orcid.org/0009-0009-8169-1520
Barbara Parol https://orcid.org/0009-0004-8956-4378
Daria Pozdniakowa https://orcid.org/0009-0001-4685-447X
Albert Synal https://orcid.org/0009-0007-3185-4374
Aleksander Sowiński https://orcid.org/0009-0008-0457-338X
Martyna Piskorz https://orcid.org/0009-0005-8542-5052
Mateusz Drążyk https://orcid.org/0009-0004-6594-7112
Mateusz Mazurek http://orcid.org/0009-0008-0066-6087
Michał Tulski https://orcid.org/0009-0001-3867-5850
Zygmunt Domagała https://orcid.org/0000-0002-2317-1932
Maciej Dobrzyński https://orcid.org/0000-0003-2368-1534
Paweł Dąbrowski https://orcid.org/0000-0001-6016-9201

References

1. Champion EM, ed. *Virtual Heritage: A Guide*. London, UK: Ubiquity Press; 2021. doi:10.5334/bck
2. Schenk T. *Introduction to Photogrammetry*. Columbus, OH: The Ohio State University; 2005.
3. Luhmann T. Close range photogrammetry for industrial applications. *ISPRS J Photogramm Remote Sens*. 2010;65(6):558–569. doi:10.1016/j.isprsjprs.2010.06.003
4. Valença J, Júlio ENBS, Araújo HJ. Applications of photogrammetry to structural assessment. *Exp Tech*. 2017;36(5):71–81. doi:10.1111/j.1747-1567.2011.00731.x
5. Mitchell HL. Applications of digital photogrammetry to medical investigations. *ISPRS J Photogramm Remote Sens*. 1995;50(3):27–36. doi:10.1016/0924-2716(95)91288-u

6. Talevi G, Pannone L, Monaco C, et al. Evaluation of photogrammetry for medical application in cardiology. *Front Bioeng Biotechnol*. 2023:11:1044647. doi:10.3389/fbioe.2023.1044647
7. Barbero-García I, Lerma JL, Mora-Navarro G. Fully automatic smartphone-based photogrammetric 3D modelling of infant's heads for cranial deformation analysis. *ISPRS J Photogramm Remote Sens*. 2020;166:268–277. doi:10.1016/j.isprsjprs.2020.06.013
8. Malian A, Azizi A, van den Heuvel FA. MEDPHOS: A new photogrammetric system for medical measurement. Paper presented at: XXth ISPRS Congress; July 12–23, 2004; Istanbul, Turkey. https://www.isprs.org/proceedings/XXXV/congress/comm5/papers/571.pdf. Accessed April 16, 2025.
9. Alshawabkeh Y, Douglas K, Matarya M, Khrisat B. Combined photogrammetric techniques and computer vision: 2D–3D recording of Gharissa, Jordan. *J Archit Conserv*. 2011;17(2):77–92. doi:10.1080/13556207.2011.10785090
10. Kihara H, Hatakeyama W, Komine F, et al. Accuracy and practicality of intraoral scanner in dentistry: A literature review. *J Prosthodont Res*. 2020;64(2):109–113. doi:10.1016/j.jpor.2019.07.010
11. Jacobs L, Dvorak J, Cornelius A, Zameroski R, No T, Schmitz T. Structured light scanning artifact-based performance Study. *Manuf Lett*. 2023;35:873–882. doi:10.1016/j.mfglet.2023.07.014
12. Bohner LOL, De Luca Canto G, Marció BS, Laganá DC, Sesma N, Tortamano Neto P. Computer-aided analysis of digital dental impressions obtained from intraoral and extraoral scanners. *J Prosthet Dent*. 2017;118(5):617–623. doi:10.1016/j.prosdent.2016.11.018
13. Burzynski JA, Firestone AR, Beck FM, Fields HW, Deguchi T. Comparison of digital intraoral scanners and alginate impressions: Time and patient satisfaction. *Am J Orthod Dentofacial Orthop*. 2018;153(4):534–541. doi:10.1016/j.ajodo.2017.08.017
14. Flügge TV, Schlager S, Nelson K, Nahles S, Metzger MC. Precision of intraoral digital dental impressions with iTero and extraoral digitization with the iTero and a model scanner. *Am J Orthod Dentofacial Orthop*. 2013;144(3):471–478. doi:10.1016/j.ajodo.2013.04.017
15. Cunha J, Fernandes GVO, Fernandes JCH, Lopes PC, Rio R. The interference of age and gender on smile characterization analyzed on six parameters: A clinical-photographic pilot study. *Medicina (Kaunas)*. 2023;59(3):595. doi:10.3390/medicina59030595
16. Silvester CM, Hillson S. A critical assessment of the potential for structure-from-motion photogrammetry to produce high fidelity 3D dental models. *Am J Phys Anthropol*. 2020;173(2):381–392. doi:10.1002/ajpa.24109
17. Stuani VT, Ferreira R, Manfredi GGP, Cardoso MV, Sant'Ana ACP. Photogrammetry as an alternative for acquiring digital dental models: A proof of concept. *Med Hypotheses*. 2019;128:43–49. doi:10.1016/j.mehy.2019.05.015
18. Zotti F, Rosolin L, Bersani M, Poscolere A, Pappalardo D, Zerman N. Digital dental models: Is photogrammetry an alternative to dental extraoral and intraoral scanners? *Dent J (Basel)*. 2022;10(2):24. doi:10.3390/dj10020024
19. Agustín-Panadero R, Peñarrocha-Oltra D, Gomar-Vercher S, Peñarrocha-Diago M. Stereophotogrammetry for recording the position of multiple implants: Technical description. *Int J Prosthodont*. 2015;28(6):631–636. doi:10.11607/ijp.4146
20. An H, Mickesh GJ, Cho D, Sorensen JA. Effect of finish line location and saliva contamination on the accuracy of crown finish line scanning. *J Prosthodont*. 2023;33(1):86–94. doi:10.1111/jopr.13658
21. Emam M, Ghanem L, Abdel Sadek HM. Effect of different intraoral scanners and post-space depths on the trueness of digital impressions. *Dent Med Probl*. 2024;61(4):577–584. doi:10.17219/dmp/162573
22. Farag EAA, Rizk A, Ashraf R, Emad Eldin F. Effect of the scanner type on the marginal gap and internal fit of two monolithic CAD/CAM esthetic crown materials: An in vitro study. *Dent Med Probl*. 2025;62(6):1131–1137. doi:10.17219/dmp/185895
23. Bolandzadeh N, Bischof W, Flores-Mir C, Boulanger P. Multimodal registration of three-dimensional maxillodental cone beam CT and photogrammetry data over time. *Dentomaxillofac Radiol*. 2013;42(2):22027087. doi:10.1259/dmfr/22027087
24. Rico JP, Brinkmann JCB, Martin MC, et al. Digital impression taking for maxillary full-arch restoration with immediate loading: A case report. *J Oral Implantol*. 2021;48(2):125–132. doi:10.1563/aaid-joi-d-20-00117
25. Castillo JC, Gianneschi G, Azer D, et al. The relationship between 3D dentofacial photogrammetry measurements and traditional cephalometric measurements. *Angle Orthod*. 2018;89(2):275–283. doi:10.2319/120317-825.1
26. Dindaroğlu F, Kutlu P, Duran GS, Görgülü S, Aslan E. Accuracy and reliability of 3D stereophotogrammetry: A comparison to direct anthropometry and 2D photogrammetry. *Angle Orthod*. 2016;86(3):487–494. doi:10.2319/041415-244.1
27. Tofangchiha M, Esfehani M, Eftetahi L, Mirzadeh M, Reda R, Testarelli L. Comparison of the pharyngeal airway in snoring and non-snoring patients based on the lateral cephalometric study: A case–control study. *Dent Med Probl*. 2023;60(1):121–126. doi:10.17219/dmp/154776
28. Sterenborg BAAM, Maal TJJ, Vreeken RD, Loomans BAC, Huysmans MCDNJM. 3D facial effects of a simulated dental build-up. *J Esthet Restor Dent*. 2016;28(6):397–404. doi:10.1111/jerd.12228
29. Page MJ, Moher D, Bossuyt PM, et al. PRISMA 2020 explanation and elaboration: Updated guidance and exemplars for reporting systematic reviews. *BMJ*. 2021:n160. doi:10.1136/bmj.n160
30. Haddaway NR, Collins AM, Coughlin D, Kirk S. The role of Google Scholar in evidence reviews and its applicability to grey literature searching. *PLoS One*. 2015;10(9):e0138237. doi:10.1371/journal.pone.0138237
31. Lefebvre C, Glanville J, Briscoe S, et al. Chapter 4: Searching for and selecting studies. In: Higgins JPT, Thomas J, Chandler J, et al., eds. *Cochrane Handbook for Systematic Reviews of Interventions version 6.5.1*. Cochrane; 2025. https://www.cochrane.org/authors/handbooks-and-manuals/handbook/current/chapter-04. Accessed April 16, 2025.
32. Whiting P, Rutjes AWS, Reitsma JB, Bossuyt PMM, Kleijnen J. The development of QUADAS: A tool for the quality assessment of studies of diagnostic accuracy included in systematic reviews. *BMC Med Res Methodol*. 2003;3(1):25. doi:10.1186/1471-2288-3-25
33. Whiting PF, Weswood ME, Rutjes AWS, Reitsma JB, Bossuyt PNM, Kleijnen J. Evaluation of QUADAS, a tool for the quality assessment of diagnostic accuracy studies. *BMC Med Res Methodol*. 2006;6(1):9. doi:10.1186/1471-2288-6-9
34. Lavorgna L, Cervino G, Fiorillo L, et al. Reliability of a virtual prosthodontic project realized through a 2D and 3D photographic acquisition: An experimental study on the accuracy of different digital systems. *Int J Environ Res Public Health*. 2019;16(24):5139. doi:10.3390/ijerph16245139
35. Brons S, van Beusichem ME, Maal TJJ, et al. Development and reproducibility of a 3D stereophotogrammetric reference frame for facial soft tissue growth of babies and young children with and without orofacial clefts. *Int J Oral Maxillofac Surg*. 2013;42(1):2–8. doi:10.1016/j.ijom.2012.07.006
36. Ceinos R, Tardivo D, Bertrand MF, Lupi-Pegurier L. Inter- and intra-operator reliability of facial and dental measurements using 3D-stereophotogrammetry. *J Esthet Restor Dent*. 2016;28(3):178–189. doi:10.1111/jerd.12194
37. Sterenborg BAAM, Maal TJJ, Vreeken R, Loomans BAC, Huysmans MCDNJM. The facial effects of tooth wear rehabilitation as measured by 3D stereophotogrammetry. *Dent J*. 2018;73:105–109. doi:10.1016/j.jdent.2018.04.014
38. Keil PU, Beuer F, Unkovskiy A, Atay E, Jennes ME. Influence of ambient light on the accuracy of different face scanning methods: An in-vitro study. *BMC Oral Health*. 2025;25(1):229. doi:10.1186/s12903-025-05594-2
39. Hartmann R, Weiherer M, Nieberle F, et al. Evaluating smartphone-based 3D imaging techniques for clinical application in oral and maxillofacial surgery: A comparative study with the Vectra M5. *Oral Maxillofac Surg*. 2025;29(1):29. doi:10.1007/s10006-024-01322-2
40. Artopoulos A, Buytaert JAN, Dirckx JJJ, Coward TJ. Comparison of the accuracy of digital stereophotogrammetry and projection moiré profilometry for three-dimensional imaging of the face. *Int J Clin Oral Maxillofac Surg*. 2014;43(5):654–662. doi:10.1016/j.ijom.2013.10.005
41. Lincoln KP, Sun AYT, Prihoda TJ, Sutton AJ. Comparative accuracy of facial models fabricated using traditional and 3D imaging techniques. *J Prosthodont*. 2016;25(3):207–215. doi:10.1111/jopr.12358
42. Majid Z, Chong A, Setan H, Ahmad AB, Rajion ZA. Natural features technique for non-contact three dimensional craniofacial anthropometry using stereophotogrammetry. *Arch Orofac Sci*. 2006;1:42–50. https://aos.usm.my/docs/Vol_1/42_50_majid.pdf. Accessed April 16, 2025.

43. Di Blasio M, Minervini G, Segù M, et al. An in-vivo study on the effect of head orientation in the measurement of anthropometric points in stereophotogrammetry. *Minerva Dent Oral Sci*. 2025;73(6):343–351. doi:10.23736/s2724-6329.24.04852-6
44. Seifert LB, Beyer M, Czok V, et al. Comparative accuracy of stationary and smartphone-based photogrammetry in oral and maxillofacial surgery: A clinical study. *J Clin Med*. 2024;13(22):6678. doi:10.3390/jcm13226678
45. Bratos M, Bergin JM, Rubenstein JE, Sorensen JA. Effect of simulated intraoral variables on the accuracy of a photogrammetric imaging technique for complete-arch implant prostheses. *J Prosthet Dent*. 2018;120(2):232–241. doi:10.1016/j.prosdent.2017.11.002
46. Ma B, Yue X, Sun Y, Peng L, Geng W. Accuracy of photogrammetry, intraoral scanning, and conventional impression techniques for complete-arch implant rehabilitation: An in vitro comparative study. *BMC Oral Health*. 2021;21(1):636. doi:10.1186/s12903-021-02005-0
47. Pinto RJ, Casado SA, Chmielewski K, Caramês JM, Marques DS. Accuracy of different digital acquisition methods in complete arch implant-supported prostheses: An in vitro study. *J Prosthet Dent*. 2024;132(1):172–177. doi:10.1016/j.prosdent.2023.07.008
48. Revilla-León M, Gómez-Polo M, Drone M, et al. Impact of scanning distance on the accuracy of a photogrammetry system. *J Dent*. 2024;142:104854. doi:10.1016/j.jdent.2024.104854
49. Rivara F, Lumetti S, Calciolari E, Toffoli A, Forlani G, Manfredi E. Photogrammetric method to measure the discrepancy between clinical and software-designed positions of implants. *J Prosthet Dent*. 2016;115(6):703–711. doi:10.1016/j.prosdent.2015.10.017
50. Tohme H, Lawand G, Chmielewska M, Makhzoume J. Comparison between stereophotogrammetric, digital, and conventional impression techniques in implant-supported fixed complete arch prostheses: An in vitro study. *J Prosthet Dent*. 2023;129(2):354–362. doi:10.1016/j.prosdent.2021.05.006
51. Tohme H, Lawand G, Eid R, Ahmed KE, Salameh Z, Makzoume J. Accuracy of implant level intraoral scanning and photogrammetry impression techniques in a complete arch with angled and parallel implants: An in vitro study. *Appl Sci*. 2021;11(21):9859. doi:10.3390/app11219859
52. Cheng J, Zhang H, Liu H, Li J, Wang HL, Tao X. Accuracy of edentulous full-arch implant impression: An in vitro comparison between conventional impression, intraoral scan with and without splinting, and photogrammetry. *Clin Oral Implants Res*. 2024;35(5):560–572. doi:10.1111/clr.14252
53. Pozzi A, Agliardi E, Lio F, Nagy K, Nardi A, Arcuri L. Accuracy of intraoral optical scan versus stereophotogrammetry for complete-arch digital implant impression: An in vitro study. *J Prosthodont Res*. 2023;68(1):172–180. doi:10.2186/jpr.jpr_d_22_00251
54. Liu M, Fu XJ, Lai HC, Shi JY, Liu BL. Accuracy of traditional open-tray impression, stereophotogrammetry, and intraoral scanning with prefabricated aids for implant-supported complete arch prostheses with different implant distributions: An in vitro study. *J Prosthet Dent*. 2024;132(3):602.e1–602.e9. doi:10.1016/j.prosdent.2024.06.006
55. Faria JC, Sampaio-Fernandes MA, Oliveira SJ, Malheiro R, Sampaio-Fernandes JC, Figueiral MH. Precision of photogrammetry and intraoral scanning in full-arch implant rehabilitation: An in vitro comparative study. *Appl Sci*. 2025;15(3):1388. doi:10.3390/app15031388
56. Estibalez-Recasens M, Serrano-Granger C, Santamaría-Laorden A, Andreu-Vázquez C, Orejas-Pérez J. Precision of stereophotogrammetry in complete arch implant recordings: A clinical study. *J Prosthet Dent*. 2025;133(5):1242–1249. doi:10.1016/j.prosdent.2024.12.021
57. Revilla-León M, Att W, Özcan M, Rubenstein J. Comparison of conventional, photogrammetry, and intraoral scanning accuracy of complete-arch implant impression procedures evaluated with a coordinate measuring machine. *J Prosthet Dent*. 2021;125(3):470–478. doi:10.1016/j.prosdent.2020.03.005
58. Revilla-León M, Gómez-Polo M, Drone M, et al. Influence of implant reference on the scanning accuracy of complete arch implant scans captured by using a photogrammetry system. *J Prosthet Dent*. 2025;133(1):252–257. doi:10.1016/j.prosdent.2024.01.008
59. Zhou Y, You L, Fan Z. Accuracy of implant abutment level digital impressions using stereophotogrammetry in edentulous jaws: An in vitro pilot study. *BMC Oral Health*. 2024;24(1):1167. doi:10.1186/s12903-024-04888-1
60. Bergin JM, Rubenstein JE, Mancl L, Brudvik JS, Raigrodski AJ. An in vitro comparison of photogrammetric and conventional complete-arch implant impression techniques. *J Prosthet Dent*. 2013;110(4):243–251. doi:10.1016/s0022-3913(13)60370-4
61. Örtorp A, Jemt T, Bäck T. Photogrammetry and conventional impressions for recording implant positions: A comparative laboratory study. *Clin Implant Dent Relat Res*. 2005;7(1):43–50. doi:10.1111/j.1708-8208.2005.tb00046.x
62. Revilla-León M, Rubenstein J, Methani MM, Piedra-Cascón W, Özcan M, Att W. Trueness and precision of complete-arch photogrammetry implant scanning assessed with a coordinate-measuring machine. *J Prosthet Dent*. 2023;129(1):160–165. doi:10.1016/j.prosdent.2021.05.019
63. Al-Khatib AR, Rajion ZA, Masudi SM, Hassan R, Townsend GC. Validity and reliability of tooth size and dental arch measurements: A stereo photogrammetric study. *Aust Orthod J*. 2012;28(1):22–29. doi:10.2478/aoj-2012-0004
64. Al-Khatib AR, Rajion ZA, Masudi SM, Hassan R, Anderson PJ, Townsend GC. Tooth size and dental arch dimensions: A stereophotogrammetric study in Southeast Asian Malays. *Orthod Craniofac Res*. 2011;14(4):243–253. doi:10.1111/j.1601-6343.2011.01529.x
65. Arapović-Savić M, Savić M, Umićević-Davidović M, Arbutina A, Nedeljković N, Glišić B. A novel method of photogrammetry measurements of study models in orthodontics. *Srp Arh Celok Lek*. 2019;147(1–2):10–16. doi:10.2298/sarh180419074a
66. Fu X, Peng C, Li Z, Liu S, Tan M, Song J. The application of multi-baseline digital close-range photogrammetry in three-dimensional imaging and measurement of dental casts. *PLoS One*. 2017;12(6):e0178858. doi:10.1371/journal.pone.0178858
67. Goellner M, Schmitt J, Karl M, Wichmann M, Holst S. Photogrammetric measurement of initial tooth displacement under tensile force. *Med Eng Phys*. 2010;32(8):883–888. doi:10.1016/j.medengphy.2010.06.004
68. Bobak CA, Barr PJ, O'Malley AJ. Estimation of an inter-rater intra-class correlation coefficient that overcomes common assumption violations in the assessment of health measurement scales. *BMC Med Res Methodol*. 2018;18(1):93. doi:10.1186/s12874-018-0550-6
69. Mahmood RSH, Hamandi SJ, Al-Mahdi AH. Creating a digital 3D model of the dental cast using structure-from-motion photogrammetry technique. *Int J Online Biomed Eng*. 2023;19(03):4–17. doi:10.3991/ijoe.v19i03.36289
70. Codari M, Pucciarelli V, Tommasi DG, Sforza C. Validation of a technique for integration of a digital dental model into stereophotogrammetric images of the face using cone-beam computed tomographic data. *Br J Oral Maxillofac Surg*. 2016;54(5):584–586. doi:10.1016/j.bjoms.2016.01.019
71. Kamel M, Elkady A, Ghanem I. Marginal fit of PMMA CAD/CAM crowns fabricated by using photogrammetry scanning method (an invitro study). *Egypt Dent J*. 2022;68(4):3537–3544. doi:10.21608/edj.2022.146431.2164
72. Mai HN, Lee DH. Effects of artificial extraoral markers on accuracy of three-dimensional dentofacial image integration: Smartphone face scan versus stereophotogrammetry. *J Pers Med*. 2022;12(3):490. doi:10.3390/jpm12030490
73. Rangel FA, Maal TJJ, Bergé SJ, et al. Integration of digital dental casts in 3-dimensional facial photographs. *Am J Orthod Dentofacial Orthop*. 2008;134(6):820–826. doi:10.1016/j.ajodo.2007.11.026
74. Yang B, Schinke J, Rastegar A, Tanyeri M, Viator JA. Cost-effective full-color 3D dental imaging based on close-range photogrammetry. *Bioengineering (Basel)*. 2023;10(11):1268. doi:10.3390/bioengineering10111268
75. Chaudhary S, Kumar S, Aggarwal R, et al. Evaluation of the accuracy of digital models generated using photogrammetry. *Cureus*. 2024;16(12):e75448. doi:10.7759/cureus.75448
76. Gašparović B, Morelato L, Lenac K, Mauša G, Zhurov A, Katić V. Comparing direct measurements and three-dimensional (3D) scans for evaluating facial soft tissue. *Sensors (Basel)*. 2023;23(5):2412. doi:10.3390/s23052412
77. Martín del Rey A, Hernández Pastora JL, Rodríguez Sánchez G. 3D medical data security protection. *Expert Syst Appl*. 2016;54(C):379–386. doi:10.1016/j.eswa.2016.02.001
78. Jankiewicz J, Mazurek M, Gajda-Janiak A, et al. Can we make anatomical specimen last forever? Technical note on photogrammetry. *Med J Cell Biol*. 2023;11(4):108–113. doi:10.2478/acb-2023-0018

79. Ebeid K, Nouh I, Ashraf Y, Cesar PF. Accuracy of different laboratory scanners for scanning of implant-supported full arch fixed prosthesis. *J Esthet Restor Dent*. 2022;34(5):843–848. doi:10.1111/jerd.12918
80. Habib S, Motiwala MA, Khan FR. Does the sagittal root position of maxillary anterior teeth affect the decision making on immediate implants in the anterior maxilla? A CBCT-based study. *Dent Med Probl*. 2024;61(1):65–70. doi:10.17219/dmp/133072
81. Clozza E. Intraoral scanning and dental photogrammetry for full-arch implant-supported prosthesis: A technique. *Clin Adv Periodontics*. 2023;14(4):244–249. doi:10.1002/cap.10269
82. Wang J, Wu YL, Ma J, Wu F, Li DH. A novel technique for implant-supported fixed complete rehabilitation based on a dynamic virtual patient. *J Dent*. 2023;137:104649. doi:10.1016/j.jdent.2023.104649
83. Pojda D, Tomaka AA, Luchowski L, Tarnawski M. Integration and application of multimodal measurement techniques: Relevance of photogrammetry to orthodontics. *Sensors (Basel)*. 2021;21(23):8026. doi:10.3390/s21238026
84. Ey-Chmielewska H, Chruściel-Nogalska M, Frączak B. Photogrammetry and its potential application in medical science on the basis of selected literature. *Adv Clin Exp Med*. 2015;24(4):737–741. doi:10.17219/acem/58951
85. Baier K, Milgrom P, Russell S, Mancl L, Yoshida T. Children's fear and behavior in private pediatric dentistry practices. *Pediatr Dent*. 2004;26(4):316–321. PMID:15344624.
86. Sieberth T, Wackrow R, Chandler JH. Motion blur disturbs – the influence of motion-blurred images in photogrammetry. *Photogramm Rec*. 2014;29(148):434–453. doi:10.1111/phor.12082
87. Shaw P, Duncan A, Vouyouka A, Ozsvath K. Radiation exposure and pregnancy. *J Vasc Surg*. 2011;53(1 Suppl):28S–34S. doi:10.1016/j.jvs.2010.05.140
88. Issa J, Kulczyk T, Rychlik M, Czajka-Jakubowska A, Olszewski R, Dyszkiewicz-Konwińska M. Artificial intelligence versus semi-automatic segmentation of the inferior alveolar canal on cone-beam computed tomography scans: A pilot study. *Dent Med Probl*. 2024;61(6):893–899. doi:10.17219/dmp/175968